



FLIR COMMERCIAL SYSTEMS, INC. - EXTECH BRAND
9 Townsend West, Nashua NH 03063 / Phone: 603.324.7800 / Fax: 603.324.7864

Declaration of Conformity

Extech Model: RH25

Description: Heat Index Psychrometer

Date of Issue: 16-Oct-13

We, FLIR Commercial Systems, Inc. - Extech Brand, 9 Townsend West, Nashua, NH 03063 declare that a sample of the product listed above has been tested by a third party for CE marking according to:

EMC Directive: 2004/108/EC
Report Number: HA110406-CE
Report Date of Issue: 9/27/2011

Standards:

EN 61326-1:2006 Class B
IEC 61326:2005
EN 55011:2009/A1:2010 Group 1 Class B
EN 61000-4-2:2009
EN 61000-4-3:2006/A2:2010
IEC 61000-4-2:2008
IEC 61000-4-3:2006+A1:2007+A2:2010

The test reports show that the product fulfills the requirement in the EC EMC Directive for CE Marking. On this basis, together with the manufacturer's own documented production control, the manufacturer (or his European authorized representative) can in his EC Declaration of Conformity verify compliance with the EC EMC Directive.

A handwritten signature in blue ink, appearing to read 'Tony Campagna', written over a horizontal line.

Tony Campagna
Director of Quality Services